


# HPLC REPORT

Structure : TB500  
Lot No : P246899-YY674348  
Column : 4.6×250mm, Welchrom C18  
Solvent A : 0.1% trifluoroacetic in 100% acetonitrile  
Solvent B : 0.1% trifluoroacetic in 100% water  
Gradient :  
                          A          B  
          0.01min  30%      70%  
          25min   55%      45%  
          25.1min 100%      0%  
          30min          STOP  
Flow rate : 1.0 mL/min  
Wavelength : 220nm  
Volume : 5ul



Peak No.	Ret Time	Height	Area	Conc.
1	11.642	7713.573	68781.523	0.6312
2	11.883	790705.000	10828960.000	99.3688
Total				100.0000